



## Thiocyanate Half Cell ISE - 1229

---

The EDT directION Thiocyanate Half Cell ion selective electrode has a solid state Crystalline membrane. It requires the use of a double junction reference electrode which has its outer chamber filled with the Ionic strength adjustment buffer (ISAB) detailed below.

The electrode is designed for the detection and analysis of Thiocyanate ions in aqueous solutions and is suitable for use in the laboratory and in on line analyzers. Please note that both the ISE and the reference electrode need to be placed in the standards or samples simultaneously and should not be more than around 10cm apart. The side filling cap of the double junction reference electrode should be open during measurements and closed during storage.

The EDT Half Cell ISE can be used immediately but pre-soaking for 5 minutes in a 100 ppm Thiocyanate solution along with the double junction reference electrode is recommended. The ionic strength of the standards and solutions should be kept constant between all standards and samples. This is achieved by the simple addition of an Ionic strength adjustment buffer (ISAB). Potassium Nitrate is ideal. A typical addition would be 1 ml of 5 molar ISAB to 100 ml of standard and sample. This solution is also used as the outer chamber reference electrode filling solution.

For low level measurements below around 50 ppm in relatively pure samples no ISAB is needed. No temperature correction is possible it is therefore important that all standards and samples should be measured at the same temperature to ensure that temperature effects are eliminated.

Begin calibration from the lowest concentration standard to avoid cross contamination. Calibration should cover the anticipated range of the samples. Rinse tips of both electrodes with de-ionised water between measurements and dab off excess water. When the response time or slope of your crystalline ISE reduces you can lightly polish the tip with very fine emery paper to refresh. To see a simple calibration please visit our website. EDT directiON produce a full range of Stock Standards and Ionic Strength Adjustment Buffers (ISABs) to save valuable time and give confidence in the quality of your results.

## SPECIFICATION TABLE

<b>Cable Length</b>	1000mm
<b>Cap Diameter</b>	16mm
<b>Concentration Range</b>	1-5800 ppm
<b>Connector</b>	BNC
<b>Diameter</b>	12mm
<b>Endpoint Time</b>	Typically 10-60 seconds
<b>Interferences</b>	Sulphide, Bromide, Cyhloride, Iodide
<b>Length</b>	155mm
<b>pH Range</b>	2-12 pH
<b>Potential Drift</b>	2mV per Day
<b>Reference Type</b>	Double Junction Reference Electrode required
<b>Resistance at 25 °C</b>	< 2.5 MOhm
<b>Temperature Range</b>	5-50 Degrees °C

For more information on this product visit [www.edt.co.uk/1229](http://www.edt.co.uk/1229)

## Related Products



Thiocyanate Ionic Strength Adjustment Buffer (ISAB) 500ml



Thiocyanate Standard Solution 1000ppm (500ml)



Double Junction Reference Electrode



Telephone +44 (0)1304 829960  
Unit 5 - Waldershare Park, Dover CT15 5DQ  
[www.edt.co.uk](http://www.edt.co.uk)

